



Package Insert

GeneChip® *Xenopus tropicalis* Genome Array

Intended Use

The GeneChip® *Xenopus tropicalis* Genome Array is a single array comprised of over 58,000 probe sets representing over 51,000 *X. tropicalis* transcripts. Sequences used in the design of the array were selected from the *Xenopus tropicalis* UniGene database (build 30, July 2006), GenBank® mRNAs for *Xenopus tropicalis* (up to September 12, 2006) and predicted gene sets from both the DOE Joint Genome Institute (Xentr4 transcripts, file creation date August 17, 2005, download date May 25, 2006) and Ensembl (release 40, download date May 2006). The GeneChip *Xenopus tropicalis* Genome Array was designed in close collaboration with the *Xenopus* research community as part of the Affymetrix® GeneChip Consortia Program.

GeneChip probe arrays are for research use only and not intended for use in diagnosis of disease. Please visit www.affymetrix.com for a complete list of supporting documentation including procedures regarding target preparation, target hybridization, fluidics station setup, probe array scan, and data analysis.

Instrumentation and Software Required

1. GeneChip® Scanner 3000 enabled for High-Resolution Scanning*
2. Affymetrix® GeneChip® Command Console® Software (AGCC)
3. GeneChip® Fluidics Station

*GeneChip Scanner 3000 High-Resolution Update is standard on all instruments shipped starting in September 2003 with serial number series 502. Previous versions, serial number series 501, will require the 00-0110 GeneChip Scanner 3000 High-Resolution Update to be installed.

Accessory Files

Fluidics

The Fluidics Script used depends on the labeling protocol, hybridization and stain reagents used. Refer to the Fluidics Script Support page to determine the appropriate fluidics protocol for this array, GeneChip® instrument system, and reagents used. The Fluidics Scripts can be downloaded from the following URL: www.affymetrix.com/support/technical/fluidics_scripts.affx

Library Files

Library files contain information about the probe array design layout and other characteristics, probe use and content and scanning and analysis parameters. These files are unique for each probe array type. The library files can be downloaded from the following URL: www.affymetrix.com/support/technical/libraryfilesmain.affx

Critical Specifications

Feature Size	11 µm
Probe Pairs/Sequence	11
Array Format	49
Hybridization Controls	<i>bioB</i> , <i>bioC</i> , <i>bioD</i> , and <i>cre</i>
Poly-A Controls	<i>dap</i> , <i>lys</i> , <i>phe</i> , <i>thr</i> , and <i>trp</i>
Housekeeping/Control Genes	<i>actin</i> , <i>ef1a</i> and <i>gapdh</i>
Hybridization Volume	200 µL. The total fill volume of the cartridge is 250 µL.
Library Files	<i>Xenopus tropicalis</i>

Ordering Information

P/N	Product Name	Description
Arrays		
901100	GeneChip® <i>Xenopus tropicalis</i> Genome Array	2 Arrays
901101	GeneChip® <i>Xenopus tropicalis</i> Genome Array	6 Arrays
901102	GeneChip® <i>Xenopus tropicalis</i> Genome Array	30 Arrays
Supporting Products		
901228 901229	GeneChip® 3' IVT Express Kit	Sufficient for 10 Reactions Sufficient for 30 Reactions
900454 900457	GeneChip® Eukaryotic Hybridizaion Control Kit	Sufficient for 30 Reactions Sufficient for 150 Reactions
900301	Control Oligo B2 (included in Hybridization Control Kit)	Sufficient for 30 Reactions
900720	GeneChip® Hybridization, Wash, and Stain Kit ¹	Sufficient for 30 Reactions

1. Each kit contains one (1) Hybridization Module, one (1) Stain Module, three (3) bottles of Wash Buffer A and one (1) bottle of Wash Buffer B, sufficient for 30 reactions. Individual kit components may be ordered separately.

Affymetrix® products can be purchased directly from Affymetrix in the United States, and many European and Asian countries. For all other territories, please view a list of our distribution partners, which can be located at: www.affymetrix.com/site/contact/index.affx.


Precautions

1. GENECHIP PROBE ARRAYS ARE FOR RESEARCH USE ONLY; NOT FOR DIAGNOSTIC PROCEDURES.
2. Avoid microbial contamination, which may cause erroneous results.
3. **WARNING: All biological specimens and materials with which they come into contact should be handled as if capable of transmitting infection and disposed of with proper precautions in accordance with federal, state, and local regulations. This includes adherence to the OSHA Bloodborne Pathogens Standard (29 CFR 1910.1030) for blood-derived and other samples governed by this act. Never pipet by mouth. Avoid specimen contact with skin and mucous membranes.**
4. **CAUTION:** Exercise standard precautions when obtaining, handling, and disposing of potentially carcinogenic reagents.
5. Exercise care to avoid cross-contamination of samples during all steps of this procedure, as this may lead to erroneous results.
6. Use powder-free gloves whenever possible to minimize introduction of powder particles into sample or probe array cartridges.

Storage, Handling and Stability

The GeneChip probe array consists of a square glass substrate mounted in a plastic cartridge (Figure 1). The glass contains an array of oligonucleotides that, when mounted, is on the inner glass surface. A chamber in the plastic housing directly under the glass acts as a reservoir where hybridization and washing occur.

Although the inner glass surface of the probe array is protected, any contamination or scratches on the outer surface of the glass can compromise the accuracy of the scan. Avoid touching the surface of the glass with your fingers. Skin oils and other substances, such as lotions or ink, can fluoresce. If the surface of the glass is noticeably dirty, it can be carefully cleaned with a non-abrasive laboratory tissue.



The GeneChip® probe arrays should be stored at 2° to 8°C. Refer to the expiration date on the package label. Do not use probe arrays or reagents after the expiration date.

Contact Information

Affymetrix, Inc.

3420 Central Expressway
Santa Clara, CA 95051 USA
E-mail: support@affymetrix.com
Tel: 1-888-362-2447 (1-888-DNA-CHIP)
Fax: 1-408-731-5441

Affymetrix UK Ltd

Voyager, Mercury Park,
Wycombe Lane, Wooburn Green,
High Wycombe HP10 0HH
United Kingdom
E-mail: supporteurope@affymetrix.com
UK and Others Tel: +44 (0) 1628 552550
France Tel: 0800919505
Germany Tel: 01803001334
Fax: +44 (0) 1628 552585

Affymetrix Japan, K. K.

ORIX Hamamatsucho Bldg, 7F
1-24-8 Hamamatsucho, Minato-ku
Tokyo 105-0013 Japan
E-mail: supportjapan@affymetrix.com
Tel: +81-3-6430-4020
Fax: +81-3-6430-4021

Please visit our web site for international distributor contact information
www.affymetrix.com

Limited License

Subject to the Affymetrix terms and conditions that govern your use of Affymetrix products, Affymetrix grants you a non-exclusive, non-transferable, non-sublicensable license to use this Affymetrix product only in accordance with the manual and written instructions provided by Affymetrix. You understand and agree that except as expressly set forth in the Affymetrix terms and conditions, that no right or license to any patent or other intellectual property owned or licensable by Affymetrix is conveyed or implied by this Affymetrix product. In particular, no right or license is conveyed or implied to use this Affymetrix product in combination with a product not provided, licensed or specifically recommended by Affymetrix for such use.

Patents

Products may be covered by one or more of the following patents: U.S. Patent Nos. 5,445,934; 5,744,305; 5,945,334; 6,140,044; 6,261,776; 6,291,183; 6,346,413; 6,399,365; 6,420,169; 6,551,817; 6,610,482; 6,733,977; 6,955,915 and D430,024 and other U.S. or foreign patents. Products are manufactured and sold under license from OGT under 5,700,637 and 6,054,270.

Trademarks

Affymetrix®, GeneChip®, NetAffx®, Command Console®, GeneTitan®, Powered by Affymetrix™, GeneChip-compatible™, Genotyping Console™, DMET™, Axiom™, and GeneAtlas™ are trademarks or registered trademarks of Affymetrix, Inc. All other trademarks are the property of their respective owners.

Copyright

© 2007-2010 Affymetrix, Inc. All rights reserved.